

CERTIFICATE OF ANALYSIS

Product Identification

Product Name Flamma® Deep Red PS Bead 0.2 um

Catalog Number PSC7001 **Appearance** Blue Liquid

Storage Conditions 4°C, protect from light

 Lot Number
 AUN00046

 Volume
 2 mL

 Solids
 2.03%

Product Analysis

Absorption * - Maximum Lot Data 645 nm

Emission * - Maximum Lot Data 665 nm

Extinction Coefficient * at 645 nm Lot Data 156,113 nm

Specification ≥ 90,000 nm

This product has been tested in a typical application and performed as expected.



Date of Issue: 13 Apr 2020

For Research Use Only

BioActs certifies on the date above that is an accurate record of the analysis of the subject lot, and that the data conform to all quality specifications.